# Quarterly Report for Goodrich Corporation Fenceline Monitoring Plan-Q1 2025

Prepared For: Goodrich Corporation 50 William White Blvd Pueblo, CO 81001

Prepared By: Montrose Air Quality Services, LLC 5270 Joyce Dr. Unit B Golden CO 80403

For Submission To: Colorado Department of Public Health and Environment 4300 Cherry Creek S Dr. Denver, Colorado 80246

Document Number: 317AA-043815-RT-941 Submittal Date: May 30, 2025



# **Table of Contents**

Executive Summary
Contact Information
Methods5
Site Description5
Instrument Description6
System Design8
Data Validation and QA/QC Procedures10
Results15
Quarterly Data Summary15
Summary of Invalidated Data18
Discussion of Invalidated Data18
Discussion of Results
Summary Plots19
Discussion of Changes to Monitoring System, Operations, and/or Procedures
Appendices
Appendix A: Calibration and QA/QC Data31
Appendix B: Qualifier Codes33
Appendix C: Field Data Sheets
Appendix D: Non-Conformance/Corrective Action Data Sheets
Appendix E: Calibration Verification Forms

# **Table of Figures**

Figure 1. The Goodrich Carbon Brake Manufacturing Process	5
Figure 2: Approximate Layout of the Open-Path Analyzers, Retroreflector Locations, and Station	d Meteorological 9
Figure 3: Timeseries of Benzene Path 1	19
Figure 4: Timeseries of H <sub>2</sub> S Path 1	19
Figure 5: Timeseries of HCN Path 1	20
Figure 6: Timeseries of Benzene Path 2	20
Figure 7: Timeseries of $H_2S$ Path 2	21
Figure 8: Timeseries of HCN Path 2	21
Figure 9: Timeseries of Benzene Path 3	22
Figure 10: Timeseries of H <sub>2</sub> S Path 3	22
Figure 11: Timeseries of HCN Path 3	23
Figure 12: Timeseries of Benzene Path 4	23
Figure 13: Timeseries of H <sub>2</sub> S Path 4	24
Figure 14: Timeseries of HCN Path 4	24
Figure 15: Timeseries of Benzene Path 5	25
Figure 16: Timeseries of H₂S Path 5	25
Figure 17: Timeseries of HCN Path 5	26
Figure 18: Timeseries of Benzene Path 6	26
Figure 19: Timeseries of H <sub>2</sub> S Path 6	27
Figure 20: Timeseries of HCN Path 6	27
Figure 21: Temperature Timeseries	28
Figure 22: Relative Humidity Timeseries	28
Figure 23: Barometric Pressure Timeseries	28
Figure 24: Wind Rose Plot	29
Figure 25: Benzene Box Plots for Paths 1 to 6	29
Figure 26: H <sub>2</sub> S Box Plots for Paths 1 to 6	

Figure 27: HCN Box Plots for Paths 1 to 6	0

# Table of Tables

Table 1: Performance Specifications for Installed Meteorological Sensors	
Table 2: Descriptions of Each Individual Path9	
Table 3: List of Automated Quality Control Parameters and Corresponding Evaluation Criteria10	
Table 4: UV DOAS QC Checks11	
Table 5: TDL QC Checks    12	
Table 6: Quarterly Data Summary15	;
Table 7: Verification Activities	
Table 8: Percent Recovery for Meteorological Parameters	
Table 9: List of Data Invalidation Codes	

## I. Goodrich Corporation Fenceline Monitoring Plan Quarterly Report- Q1 2025

### II. Executive Summary

This report summarizes the findings related to the Goodrich Corporation fenceline monitoring plan during the period of Januaty 1<sup>st</sup> of 2025 to March 31<sup>st</sup> of 2025 (Q1 of 2025). The data collected during this period were validated following all procedures described in the Goodrich Corporation fenceline monitoring plan. This report includes tables with the validated and invalidated data, statistical analysis results and timeseries of the compounds of interest and meteorological parameters.

### III. Contact Information

For any questions related to this report please contact: -David Andreaccio (<u>David.Andreaccio@collins.com</u>) and -Michael Young (<u>Michael.Young@collins.com</u>)

### IV. Methods

### A. Site Description

Goodrich operates an aircraft brakes manufacturing facility at 50 William White Blvd, Pueblo, CO 81001. The carbon brake manufacturing process is a series of seven operational steps: The textile of preforms, carbonization of preforms, chemical vapor deposition (CVD) of preforms, intermediate machining of preforms, an additional CVD cycle, final dimensional machining, and final assembly.



Figure 1. The Goodrich Carbon Brake Manufacturing Process

These specific processes can be more generally separated into four manufacturing areas; textile, furnace operations, machining, and finishing. The textile process transforms raw polyacrlyonitrile (PAN) fibers into a three-dimensional matrix or brake preform. Brake preforms are then batch processed in high-temperature, low-pressure reactors (furnaces). The two major processes completed are carbonization and densification. Carbonization converts the raw PAN preforms to a carbon fiber preform and removes impurities. The densification process cracks a feed hydrocarbon stream to infiltrate and sequester molecular carbon on the carbon fiber preform. The machining of preforms is completed at two stages; once in the middle of the densification and once following densification. After the final machining operations, the

final assembly operations include application of an oxidation protection system and hardware installation. Additionally, the carbon brake manufacturing process requires extensive chemical process infrastructure including waste heat recovery, steam generation, cooling water systems, gas purification, and gas storage and delivery systems

## **B.** Instrument Description

### 1. Open-Path Monitors

The Goodrich Corporation fenceline air monitoring system includes both open-path tunable diode laser spectrometers (TDLAS), and open-path ultraviolet Doppler optical absorption spectrometers (UVDOAS). Open-path monitors operate by projecting a beam of light through open air to retroreflectors that reflect the light back to the monitor where spectral absorption characteristics are measured. As the light travels along the path length a certain amount of this light will be absorbed by the various chemical species present in the air. Because all gases absorb light differently according to their own unique spectral characteristics, it is possible to use measurements of absorption intensity at specific wavelengths as a proxy for measuring a target gas' concentration in the air.

Therefore, along a known path length, an absorption measurement taken at the appropriate wavelength for the target molecule can easily be used to solve for its average concentration over the length of the beam.

The Goodrich Corporation system will consist of three TDL analyzers and six UVDOAS analyzers at the locations shown in Figure 2 and as outlined in Table 2. The light is transmitted to a retroreflector and back to a detector co-located with the transmitter. The analyzer software will provide five-minute and hourly-average concentration measurements for each path.

### - Open Path (OP) Ultra Violet Differential Optical Absorption Spectroscopy (UVDOAS)

For the monitoring of benzene, the Goodrich Corporation facility uses Open Path (OP) Ultraviolet Differential Optical Absorption Spectroscopy (UVDOAS). This technology quantifies concentrations of gaseous compounds by measuring the absorption of ultraviolet light by chemical compounds in the air and applying the Beer-Lambert Law. UVDOAS typically uses unique absorptions of specific wavelengths of ultraviolet light in a wavelength range of 245 to 380 nanometers (nm). Benzene peaks are found close to the 253 nm wavelength.

Open path UVDOAS instrumentation consists of a light source, transmitting and receiving optics (telescopes), a spectrometer, a reflector or receiver, a detector, and a data processing computer. A Xenon light source provides light, which is focused in a collimated beam before it is sent through a transmitting telescope and into the measurement path. A receiving telescope collects the light and directs it to the spectrometer which diffracts the light onto the detector. The detector is typically a solid-state array such as a charge-coupled device (CCD). This allows the detector to collect light of different wavelengths without moving parts. The spectra bands can be extracted from the spectrum and compared to reference spectra to determine which compounds were present along the path and at what concentrations.

A combination of monostatic and bistatic open path instruments have been selected to reduce the need for substantial power at the retroreflector sites and improve detection limits by increasing effective path lengths.

The Goodrich Corporation facility uses the UV Sentry Open Path Multi-Gas Analyzer (UV Sentry) manufactured by Cerex Monitoring Solutions, LLC for the monitoring of benzene. The UV Sentry uses no moving parts to wear out, it should not fail or require calibration, which keeps consumables and maintenance to a minimum. The UV Sentry has an on-board computer and saves raw spectral data independent of calibration. These spectra may be used at any time to verify real time measurements.

Additionally, the UV Sentry records signal intensity and minimum detection limits (MDLs) for benzene in real time as data quality indicators. Real time MDL output supports both American Society for Testing and Materials (ASTM) and USEPA methods. The UV Sentry also has a flow through calibration cell to allow for regular QA audits and bump tests.

### - Open Path (OP) Tunable Diode Laser Absorption Spectroscopy (TDLAS)

For the monitoring of Hydrogen Sulfide and Hydrogen Cyanide<sup>1</sup>, an Open Path (OP) Tunable Diode Laser Absorption Spectroscopy (TDLAS) is used. OP-TDLAS offers some significant operational and cost advantages over other measurement technologies such as Fourier Transform Infrared Spectroscopy (FTIR). Tunable diode lasers (TDL) are designed to focus on single absorption wavelengths specific to a compound of concern in the gaseous form. They are capable of achieving low detection limits and are generally interferent-free. Similar to UVDOAS, quantitative measurements in direct gas phase laser absorption spectroscopy are based on the Beer-Lambert Law. A TDL uses a diode to generate light within a narrow frequency range that contains a relatively unique absorption wavelength of the chemical of interest. The laser frequency is "tuned" by changing the temperature of the diode or the current being fed to the diode or both so that it matches the spectral absorption line of interest.

Similar to the UVDOAS system, the OP-TDLAS system consists of a light source, a spectrometer, a reflector, a photodiode detector, and a data processing computer. Monostatic (as opposed to bistatic) open path instruments have been selected to reduce the need for substantial power at the retroreflector sites, and improve detection limits by increasing effective path lengths.

The Goodrich Corporation facility uses the LasIR<sup>™</sup> Fence Line Monitoring Gas Analyzer manufactured by Unisearch Associates Inc. for the monitoring of Hydrogen Sulfide and Hydrogen Cyanide.<sup>1</sup> The LasIR<sup>™</sup> allows one laser to send beams at two different wavelengths down each path length (one for each compound). Additionally, the beam can be split allowing it to monitor two path lengths with one laser. The controller uses a near infrared (NIR) Tunable Diode Laser Absorption Spectrometer System utilizing a single mode laser mounted in a thermoelectric cooler. A Windows based software package displays the data on a host laptop PC. The LasIR<sup>™</sup> also has a flow through calibration cell to allow for regular QA audits and bump tests.

### 2. Meteorological Monitors

The meteorological monitoring tower is located at the northwest end of the Goodrich Corporation property. This tower is outfitted with high quality meteorological instruments, as outlined in Table 1, and are capable of making accurate real time measurements continuously. All sensors will be connected to a datalogger which will store the data, as well as broadcast it out to a cellular modem so that data can be viewed or downloaded at any time, from anywhere. The specific meteorological instruments chosen meet EPA specifications for accuracy, range and resolution (Table 1) and have been deemed appropriate for use in the fenceline monitoring system. Data from these sensors will be used to calculate 1-hour rolling averages updated every five minutes.

<sup>&</sup>lt;sup>1</sup> Hydrogen sulfide is neither used nor stored at, nor are they emitted from the Goodrich Corporation. Therefore, the facility does not have the potential to emit this compound, which comprise "Covered Air Toxics" under HB21-1189.

Parameter	Sensor Make and model	Reporting units	Accuracy	Range	
Horizontal wind speed	Met One 010C	Meters per second (m/s)	± 0.1	0 to 55	
Horizontal wind direction	l wind on Met One 020D Degrees (°)		± 3	0 to 360	
Temperature	Met One 065	Degrees of Celsius (°C)	± 0.15	-30 to +50	
Relative humidity	Met One 083F/0/35	Percentage (%)	± 2	0 to 100	
Barometric pressure	Met One 0192	Atmospheres (atm)	± 0.001	0.3 to 1.09	

Table 1: Performance Specifications for Installed Meteorological Sensors

### C. System Design

The fenceline monitoring system will utilize three primary shelters to house the open path analyzers. Shelters 1 and 6 (Figure 2) will house one (1) monostatic open-path tunable diode laser (TDL) analyzer, one (1) monostatic open-path ultraviolet differential optical absorption (UV-DOAS) analyzer and one (1) bistatic open-path UV-DOAS receiver. Shelter 3 will house one (1) monostatic open-path TDL analyzer, and two (2) bistatic open-path UV-DOAS receivers.

Each open-path analyzer location will have multiple paths, where each path is measured continuously. At the end of each path there will be a retroreflector opposite the analyzer for the monostatic systems and a receiver opposite the analyzer for the bistatic systems. Each path ID consists of a number. Path numbers range from 1 to 6 (Figure 2, Table 2). The specific locations for all open path equipment were selected in order to provide coverage of all facility emission sources within the constraints of the facility footprint.

This monitoring program also includes meteorological monitoring as required in HB21-1189. Meteorological monitoring is necessary to characterize wind patterns for understanding movement of the three target compounds and potential sources of emissions, whether they originate from the site or a neighboring facility. A meteorological tower will be installed near Shelter 6 (Figure 2) so that power can be shared.



Figure 2. Approximate Layout of the Goodrich Corporation FLMP

Table 2: Descriptions of Each Individual Path

Path	Path Length (one way)	Compounds				
		Hydrogen sulfide				
1	558 meters	Hydrogen cyanide				
		Benzene				
		Hydrogen sulfide				
2	283 meters	Hydrogen cyanide				
		Benzene				
		Hydrogen sulfide				
3	613 meters	Hydrogen cyanide				
		Benzene				
		Hydrogen sulfide				
4	566 meters	Hydrogen cyanide				
		Benzene				
		Hydrogen sulfide				
5	297 meters	Hydrogen cyanide				
		Benzene				
		Hydrogen sulfide				
6	569 meters	Hydrogen cyanide				
		Benzene				

## D. Data Validation and QA/QC Procedures

### -Automated Quality Control Procedures

Many Quality Control procedures for the fenceline monitoring network are integrated directly into the AirSense data platform and are outlined as follows. These automated procedures allow for the ability to screen data not suitable for public display due to atmospheric or operational issues. These automated quality control checks include:

- Inspection of daily reports generated by the AirSense platform which summarize data recovery for each analyzer/sensor and suspect data flags;
- Monitoring of real time alerts and daily reports generated by the AirSense data platform that flag:
  - o No data
  - Data sticking if values are repeated for a number of sampling intervals (does not apply to data below the detection limit)
  - Range exceedances if values are outside a reasonable minimum or maximum value
  - o Data recovery
  - Monitoring instrument parameters that may indicate equipment degradation / failure or a need for maintenance and / or cleaning
  - Signal intensity (open path instruments)
  - Instrument or sensor alarms or error codes
  - Analyzer and shelter temperatures
  - Laser parameters (TDL instruments)

#### Table 3: List of Automated Quality Control Parameters and Corresponding Evaluation Criteria

Instrument	Automated Quality Control Parameter	Definition	Units	Evaluation criteria
	MDL	Minimum detection limit	PPB	< 25% of alert threshold
UV-DOAS	R <sup>2</sup>	Percentage peak match	%	> 64
	Signal intensity	Signal intensity at full scale	%	> 40
	UV spectrometer temperature		٥C	35
TDL	MDL	Minimum detection limit	PPB	< 25% of alert threshold for paths 1,3,4,6 <50% of alert threshold

				for H2S paths 2 and 4
TDL	Absolute Signal	Detector Signal	mA	> 0.1
	Laser temperature stability	Absolute value of (laser temperature- laser temperature in long average) *100/ laser temperature in long average	%	< 5
	R	Peak correlation		> 0.8

#### -Instrument Quality Control Checks

Both the UV-DOAS and TDL systems are designed to require only modest service and maintenance. Section 5.4 of the FLMP summarizes the UV-DOAS and TDL maintenance activities as recommended by the manufacturer. These activities will help ensure data integrity and maximize up-time. For the UV-DOAS system, a calibration verification bump test is performed on a quarterly basis using a flow through cell. For the UV-DOAS system, precision is calculated by evaluating the variance of pollutant concentrations during a period of low atmospheric variability. Five-minute data are selected when concentrations are well above the minimum detection limit (MDL) during periods of low variability. The precision can then be determined by calculating the coefficient of variation (CV). For the UV-DOAS, robustness can be determined by calculating the desired signal intensity in order for the benzene minimum detection limit to be lower than 25% of the notification threshold. If the measured signal intensity is found to be below the desired value, corrective action will be required (e.g., replace light source, instrument alignment, etc.). The QC checks for the UVDOAS are summarized in Table 4.

QA/QC Check	Frequency	Acceptance Criteria		
Accuracy and precision (Bump	Quarterly	Accuracy: ≤ 30% of reference gas value		
Test)		Precision: ± 25%		
Baseline Stability	Continuous	± 5%		
Signal intensity	Continuous	>60%		
Robustness	Continuous	Compound MDL lower than 25% of notification threshold		

For the TDL system, a calibration verification bump test is performed on a quarterly basis. The bump test simulates system-observed gas content at the required path average concentration and is used to verify

that the system can detect concentrations at or below the levels of concern. For the TDL system, precision will be calculated by evaluating the variance of pollutant concentrations during a period of low atmospheric variability. Five-minute data will be selected when concentrations are well above the minimum detection limit during periods of low variability. The precision can then be determined by calculating the coefficient of variation (CV). If there are no periods of low variability with concentrations above the minimum detection limit, bump test data will be used for the precision determination. For the TDL system, robustness can be determined by calculating the desired signal intensity for the hydrogen sulfide and hydrogen cyanide minimum detection limit to be lower than 25% (and 50% for H2S Paths 2 and 5) of the corresponding notification thresholds. If the measured signal intensity is found to be below the desired value, corrective action will be required (e.g., replace laser, instrument alignment, etc.). The QC checks for the TDL are summarized in the table as follows.

QA/QC Check	Frequency	Acceptance Criteria			
Accuracy and precision (Bump	Quarterly	Accuracy: ≤ 30% of reference gas value			
Test)		Precision: ± 25%			
Baseline Stability	Continuous	± 5%			
Signal intensity (Absolute Power)	Continuous	>0.1 mA			
Robustness	Continuous	Compound MDL < 25% of alert threshold for paths 1,3,4,6 and Compound MDL <50% of alert threshold for H2S paths 2 and 4			

#### Table 5: TDL QC Checks

Wind speed, wind direction, temperature, relative humidity and barometric pressure measurement systems will be aligned, tested and calibrated at the time of installation and at six-month intervals thereafter using test equipment traceable to NIST or other authoritative standards and following standard operating procedures. Calibrations are performed immediately following scheduled semi-annual meteorological audits and performance of scheduled preventive and/or corrective maintenance for the monitoring instruments. Following initial startup calibrations and continuing throughout the monitoring program, the field operator performs quarterly site checks on the meteorological monitoring systems. In the course of these checks, sensors will be observed for proper operation. The monitoring instruments and support equipment are visually inspected to confirm operational integrity. The current data logger readings are assessed for agreement with prevailing conditions.

#### -Data Quality Assurance

All continuous data from the monitoring equipment are transferred to the cloud-based servers every five minutes. Each business day, a data technician checks the data files to ensure that all data were successfully transmitted and stored in the database. If data are missing, they are manually retrieved from the computers that control each piece of equipment or the on-site data logger for the meteorological

equipment. This data is the raw data collected from the instrument computers or data logger and is considered "Level 1" data. These data are used to monitor instrument operations on a regular basis but are not used for reporting until subject to further review and validation. Level 1 (raw) data files are kept intact and unedited. These data are not subject to reduction or reformatting.

"Level 1" data are "raw" data; i.e., data obtained directly from the instrument computers or data logger that have not yet been subjected to quality assurance review. Electronic files of the raw data record are archived "as is"; no alteration is made to the raw data files. All data processing, editing and validation work is accomplished by working with copies of the raw data files produced by the data management system software upon request. Level 1 data are manually reviewed for reasonableness and completeness. Initial (daily) review of the data occurs no more than four days after sample acquisition because of weekends and holidays. Daily data review includes checking for status or event flags, reasonableness of reported averaged data values (out-of-range, inconsistent or excessive transition values) and any missing data periods. The operating status of each instrument is also reviewed (e.g., sample flow rates; other internal operating parameters). Meteorological data are reviewed for agreement with local seasonal and prevailing conditions and internal consistency. These daily reviews support "Level 2" validation of the data and provide a decision basis for investigative actions, instrument adjustment and calibration. The data analyst annotates the separate data processing file (i.e., an electronic copy of the original raw data file) and produces a summary report of any suspect data or out-of-tolerance operating conditions. Any situation requiring investigative and/or corrective action is immediately brought to the attention of the Project Manager and Technical Lead. A "Non-Conformance / Corrective Action" (NC/CA) report documenting all pertinent information regarding suspect data, a non-conformance event or out-of-tolerance operating condition is generated and updated with further information as it becomes available until the problem is fully resolved.

All data reporting forms and activity logs completed during the previous month are stored in Montrose's local Denver office and are reviewed against the electronic data record on a monthly basis in support of data processing and validation. Monthly review of the field monitoring documentation will include:

- All completed routine site check form
- Documentation of the QC tests performed on the monitors during the previous month
- Documentation of any maintenance activities performed on the monitors during the previous month
- Documentation of any quality assurance audits performed on the meteorological sensors during the previous month
- Documentation of any Non-Conformance/Corrective Action (NC/CA) events that occurred during the previous month

During "Level 2" data validation, the data file of each continuously-monitored parameter is processed at monthly intervals to develop an initial data report to be reviewed for completeness and correctness. Any corrections or additions to the raw "Level 1" data file are annotated in the processing data file with explanatory comments. Any hours incorporating a test, calibration or other quality control check, corrective or preventive maintenance, instrument malfunction, power failures, weather event, etc. are removed from the data set and annotated with the appropriate null data code (for detail on null data codes and corresponding descriptions see Table 11 of Appendix F). Results of this review, including any data losses equal to or greater than one hourly block average, are documented and dated by the data technician in "Level 2" data files. The data technician enters and annotates any null data codes or corrections required in the "Level 2" electronic data file. When all entries or corrections are complete, the data are designated as "Level 2 - Final" data, and are archived for subsequent final data validation review.

"Level 3" data validation review is performed by senior project personnel other than the data processing analyst. During the Level 3 data validation process, data losses due to activity or instrument malfunction are corroborated against documentation noted by the station field operators on completed field forms. The field form record identifying data affected by these activities and events are inter-compared with corresponding status flags entered by the operator in the digital data record. Documented results of QA/QC checks performed on each analyzer are evaluated with respect to relevant acceptance and performance criteria outlined in the fenceline monitoring plan. Reports documenting unacceptable operating conditions or non-conformance/corrective action (NC/CA) events that may have adversely impacted data quality are also reviewed. If discrepancies or questionable data values are identified during the validation process, the entire data record is reviewed (including all annotated corrections made for Level 2 data). Any additional corrections or revisions made to the data report file during the data validation review are documented, dated and signed by the validation reviewer. The corrections are then entered into the electronic data file and reprocessed. A separate file containing the corrections is checked for accuracy against the documented corrections. When all corrections are complete and checked, a final "Level 3 - Validated" data file is produced.

## V. Results

# A. Quarterly Data Summary

 Table 6: Quarterly Data Summary

Path	Compound	Number of Exceedances <sup>1</sup>	0th <sup>2</sup>	25th <sup>2</sup>	50th <sup>2</sup>	75th <sup>2</sup>	100th <sup>2</sup>	Avg	Pct Detect <sup>3</sup>	Pct Valid <sup>₄</sup>	Median 1hr DL⁵
1	Benzene	0	0.4	1.0	2.0	5.5	246.0	5.1	0.00%	97.6%	1.9
1	H2S	0	0.0	3.1	8.4	16.9	56.1	11.5	0.10%	91.0%	8.6
1	HCN	0	0.0	0.3	0.8	1.4	6.2	0.9	0.79%	78.3%	0.5
2	Benzene	0	0.3	1.1	1.5	2.0	497.0	3.4	0.04%	69.8%	2.1
2	H2S	0	0.0	8.0	12.9	20.6	242.1	17.7	0.10%	85.2%	17.5

2	HCN	0	0.1	0.4	1.1	2.5	98.5	2.1	2.00%	69.7%	1.6
3	Benzene	0	0.1	0.3	0.4	0.7	90.9	0.8	0.10%	98.2%	0.5
3	H2S	0	0.5	3.2	4.7	7.1	30.9	5.6	0.00%	83.8%	6.2
3	HCN	0	0.0	0.6	0.9	13	6.9	11	1 70%	63.4%	1.0
4	Benzene	0	0.2	0.7	13	11 7	396.0	83	0%	96.7%	16
4	H2S	0	0.1	3.1	5.2	9.1	48.8	7.2	0.00%	82.3%	6.2
4	HCN	0	0.1	0.9	1.8	2.7	12.6	2.0	5.20%	62.8%	1.1

5	Benzene	0	0.2	0.8	1.4	2.3	592.0	2.7	0.00%	95.9%	1.7
5	H2S	0	0.2	8.9	15.5	23.7	111.5	18.4	0.07%	79.2%	19.5
5	HCN	0	0.1	0.6	1.4	2.6	100.1	2.1	0.50%	78.8%	1.6
6	Benzene	0	0.1	0.6	0.9	16	260.3	2.0	0.00%	97.1%	12
6	H2S	0	0.1	3.7	5.7	8.6	45.2	7.0	0.00%	77.9%	7.2
6	HCN	0	0.0	0.1	0.4	0.7	5.8	0.5	0.45%	76.3%	0.4

<sup>1</sup> number of 1-hour measurements above the notification threshold value

 $^{2}$  data quartiles = the value at which a defined percentage of data existing below this value (valid data only)

<sup>3</sup> the percentage of hourly averages above the detection limit (DL) as compared to the total possible hourly averages (excluding data collected during QA/QC activities, calibration, or maintenance).

<sup>4</sup> the proportion of the 1h measurements that pass all data verification measures compared to the possible hourly averages.

<sup>5</sup> the median 1-hr detection limit observed across validated measurements per compound for the month specified.

## B. Summary of Invalidated Data

The invalidated data can be found in file "Goodrich Corporation FLMP Data Packet\_Q1 2025". All 5min data have been validated based on the procedures described in the Goodrich Corporation fenceline monitoring plan.

## C. Discussion of Invalidated Data

The data was validated based on the procedures mentioned in the fenceline monitoring plan. During this first quarter of the fenceline monitoring program operation, there was a relatively low invalidation rate for benzene, H2S and HCN with an average valid data percentage of approximately 82.4%. For H2S, some lower percentage rates were observed in Paths 2 and 4. The reason was related to these paths being shorter compared to the rest and therefore having higher detection limits causing the analyzers not to be able to meet the criteria for the detection limits (minimum detection limit was more than 50% of threshold). Goodrich is not emitting nor storing H2S but there is a nearby H2S source related to a pumping station.

### D. Discussion of Results

As shown in the summary plots, the concentration of the three compounds of interest was below detection limit in most cases. There were no threshold exceedances during Q1 of 2025 for any of the compounds. For benzene the average median MDL value was around 4.5 ppb, for H2S the average median MDL value was approximately 4.9 ppb, and for HCN the corresponding average median MDL was around 4.4 ppb. As discussed in Section C, the higher H2S MDL values for Paths 2 and 4 are related to the interferences of this compounds with water and CO2 which can cause increased signal noise levels combined with the shorter path length. Goodrich Corporation does not store nor emit H2S.



Figure 3. Timeseries of Benzene Path 1



Figure 4. Timeseries of H2S Path 1





Figure 5. Timeseries of HCN Path 1

Figure 6. Timeseries of Benzene Path 2







Figure 8. Timeseries of HCN Path 2







Figure 10. Timeseries of H2S Path 3



Figure 11. Timeseries of HCN Path 3



Figure 12. Timeseries of Benzene Path 4



Figure 13. Timeseries of H2S Path 4



Figure 14. Timeseries of HCN Path 4



Figure 15. Timeseries of Benzene Path 5



Figure 16. Timeseries of H2S Path 5



Figure 17. Timeseries of HCN Path 5



Figure 18. Timeseries of Benzene Path 6



Figure 19. Timeseries of H2S Path 6



Figure 20. Timeseries of HCN Path 6







Figure 22. Relative Humidity Timeseries (2025)



Figure 23. Barometric Pressure Timeseries (2025)







Figure 25. Benzene Box Pots for Paths 1 to 6.



Figure 26. H<sub>2</sub>S Box Plots for Paths 1 to 6.

![](_page_29_Figure_2.jpeg)

Figure 27. HCN Box Plots for Paths 1 to 6.

### F. Discussion of Changes to Monitoring System, Operations and/or

### Procedures

Three main changes were performed to the fenceline monitoring plan procedures which are related to the automated QA/QC checks.

- 1. <u>UV spectrometer temperature</u>: the UV spectrometers were calibrated by the manufacturer at 35°C instead of the 39°C that the older models were used to be calibrated at. Thus, for the automated QA/QC checks, we changed the acceptance criteria to accommodate the updated spectrometer calibration conditions.
- 2. <u>TDL signal intensity</u>: the manufacturer recommended to monitor the absolute detector power instead of the signal intensity. The reason was related to the fact that the laser signal intensity is affected by multiple instrument parameters (I/O Gain, Signal Gain, signal collimation etc.). Due to these interferences, the signal intensity values that are reported by the analyzer could potentially not be representative of the actual signal power that is measured by the detector. To avoid these issues, we replaced the "signa intensity" parameter on the automated QA/QC checks with the parameter "absolute detector power". The criteria for the data to be considered valid is the absolute detector power to be >0.1 mA.
- 3. <u>H2S MDL</u>: For the purposes of the data validation the H2S MDL threshold was increased from 25% to 50% of the alert threshold for Paths 2 and 5. This change was related to the fact that these two paths are shorter compared to the rest which causes them to have higher MDL values. The fenceline monitoring plan has been updated to reflect those changes and has been submitted to the Division for review.

### VI. Appendices

### A. Appendix A: Calibration and QA/QC Data

Date	Type of Verification	Path	Path Length <sup>1</sup>	Analyzer	Compound	Expected Concentration	Measured Concentration	Accuracy (%)	Precision (%)
3/5/2025	Bump test	1	588	UVDOAS	Benzene	100	115	15.2	6.6
3/5/2025	Bump test	1	588	UVDOAS	Benzene	200	195	5.7	6.4
3/5/2025	Bump test	2	566	UVDOAS	Benzene	100	98	6	7.1
3/5/2025	Bump test	2	566	UVDOAS	Benzene	200	230	15.2	2.2
3/5/2025	Bump test	3	613	UVDOAS	Benzene	100	88.4	11.6	5.7
3/5/2025	Bump test	3	613	UVDOAS	Benzene	200	207	5.7	5.4
12/30/2024	Bump test	4	566	UVDOAS	Benzene	100	123	22.6	6.3
3/5/2025	Bump test	4	566	UVDOAS	Benzene	200	199	6.6	9.7
3/5/2025	Bump test	5	594	UVDOAS	Benzene	100	118	18.2	14.8
3/5/2025	Bump test	5	594	UVDOAS	Benzene	200	209	4.7	3.3

**Table 7. Verification Activities** 

3/5/2025	Bump test	6	569	UVDOAS	Benzene	100	115	15	12.3
3/5/2025	Bump test	6	569	UVDOAS	Benzene	200	215	12.2	13.8
3/5/2025	Audit Module	1	1116	TDL	H2S	500 ppmm	543	8.6	3.1
3/5/2025	Audit Module	1	1116	TDL	H2S	625 ppmm	636	2.9	4.4
3/5/2025	Audit Module	2	566	TDL	H2S	500 ppmm	487	2.6	1.9
3/5/2025	Audit Module	2	566	TDL	H2S	625 ppmm	688	10.8	6.8
3/5/2025	Audit Module	3	1226	TDL	H2S	500 ppmm	475	5	1.5
3/5/2025	Audit Module	3	1226	TDL	H2S	625 ppmm	582	6.9	1
3/5/2025	Audit Module	4	1132	TDL	H2S	500 ppmm	401	19.8	2.1
3/5/2025	Audit Module	4	1132	TDL	H2S	625 ppmm	556	11	1.3
3/5/2025	Audit Module	5	594	TDL	H2S	500 ppmm	516	3.3	2.2
3/5/2025	Audit Module	5	594	TDL	H2S	625 ppmm	711	13.7	1.3
3/5/2025	Audit Module	6	1138	TDL	H2S	500 ppmm	521	4.2	1.1
3/5/2025	Audit Module	6	1138	TDL	H2S	625 ppmm	689	10.2	2.4
3/5/2025	Audit Module	1	1116	TDL	HCN	1010 ppmm	1054	4.3	0.2
3/5/2025	Audit Module	1	1116	TDL	HCN	420 ppmm	494	17.7	0.6
3/5/2025	Audit Module	2	566	TDL	HCN	1010 ppmm	1054	4.4	0
3/5/2025	Audit Module	2	566	TDL	HCN	420 ppmm	491	17	0.3
3/5/2025	Audit Module	3	1226	TDL	HCN	1010 ppmm	1041	3	0.1
3/5/2025	Audit Module	3	1226	TDL	HCN	420 ppmm	482	14.7	0.4
3/5/2025	Audit Module	4	1132	TDL	HCN	1010 ppmm	1037	2.7	0.2
3/5/2025	Audit Module	4	1132	TDL	HCN	420 ppmm	401	4.6	2.5
3/5/2025	Audit Module	5	594	TDL	HCN	1010 ppmm	1034	2.4	0.1
3/5/2025	Audit Module	5	594	TDL	HCN	420 ppmm	471	12.1	0.3
3/5/2025	Audit Module	6	1138	TDL	HCN	1010 ppmm	1042	3.2	0.1
3/5/2025	Audit Module	6	1138	TDL	HCN	420 ppmm	483	15	0.3

<sup>1</sup>path length in meters

# Table 8: Percent Recovery for Meteorological Parameters

Parameter	Percent Data Recovery
Wind Speed	100%
Wind Direction	100%
Temperature	100%
Humidity	100%
Pressure	100%

Qualifier Code	AQS Definition *(additional information added in parentheses)	Type or Related Action
AB	Technician Unavailable. *(use if this affects scheduled QA/QC or necessary maintenance)	Null Data Qualifier
AD	Shelter Storm Damage.	Null Data Qualifier
AG	Sample Time out of Limits. *(e.g., use if integration time is out of manufacturer recommended range and signal intensity and MDL cannot meet the critical criteria mentioned in the FLMP)	Null Data Qualifier
AI	Insufficient Data. (cannot calculate)	Null Data Qualifier
AL	Voided by Operator. *(e.g., Datum rejected by data validators)	Null Data Qualifier
AM	Miscellaneous Void.	Null Data Qualifier
AN	Machine Malfunction *(can be used for issues such as an instrument being out of alignment, or an analyzer being offline due to connection problems or instrument failure)	Null Data Qualifier
AO	Bad Weather. *(Use if weather impacts open-path instrument operation/function)	Null Data Qualifier
AP	Vandalism. *(Use if vandalism impacts open-path instrument operation/function)	Null Data Qualifier
AQ	Collection Error. *(use specifically for low analyzer signal events, or when a low analyzer signal prevents the reported data from meeting the critical criteria, while the calculated MDL is lower than 25% of notification threshold)	Null Data Qualifier
AT	Calibration.	Null Data Qualifier
AU	Monitoring Waived.	Null Data Qualifier
AV	Power Failure.	Null Data Qualifier
AW	Wildlife Damage. *(Use if damage impacts open-path instrument operation/function)	Null Data Qualifier
AX	Precision Check.	Null Data Qualifier
AY	QC Control Points (zero/span).	Null Data Qualifier
AZ	QC Audit.	Null Data Qualifier
BA	Maintenance/Routine Repairs.	Null Data Qualifier
BH	Interference/co-elution/misidentification.	Null Data Qualifier
BJ	Operator Error.	Null Data Qualifier
BK	Site computer/data logger down.	Null Data Qualifier
BL	QA Audit.	Null Data Qualifier
BM	Accuracy check.	Null Data Qualifier
DA	Aberrant Data (Corrupt Files, Spikes, Shifts).	Null Data Qualifier
DL	Detection Limit Analyses.	Null Data Qualifier

EC	Exceeds Critical Criteria. *(use when data exceeds critical criteria, such as for MDL)	Null Data Qualifier
IA	African Dust. *(use for any dust event)	Informational
IT	Wildfire-U.S. *(use for any wildfire event)	Informational
J	Construction/Repairs in Area.	Informational
LJ	Identification of Analyte Is Acceptable; Reported Value Is An Estimate.	Quality Assurance Qualifier
MD	Value less than MDL.	Quality Assurance Qualifier
NS	Influenced by nearby sources. *(e.g., in the event of emissions influenced by nearby sources)	Quality Assurance Qualifier
QP	Pressure Sensor Questionable. *(e.g., use if cell pressure is out of range, indicating malfunction)	Quality Assurance Qualifier
QT	Temperature Sensor Questionable. *(e.g., use if cell temperature is out of range, indicating malfunction)	Quality Assurance Qualifier
QV	Quality Control Multi-point Verification.	Null Data Qualifier
QX	Does not meet QC criteria. *(e.g., data exceeds automatic criteria for rejection)	Quality Assurance Qualifier
SC	Sampler Contamination.	Null Data Qualifier
ST	Calibration Verification Standard.	Null Data Qualifier
тс	Component Check & Retention Time Standard. *(use this code for additional instrument checks, e.g., a robustness tests)	Null Data Qualifier

# C. Appendix C: Field Data Sheets

11:16 AM 1/8/2025 Montrose Remote KL Bad weather conditions 1/6-1/8

12:13 PM 1/25/2025 Montrose Remote KL Bad weather conditions 1/25-1/26

2:15 PM 1/8/25 Montrose onsite CF Aligned path 1 & 2 UVs, path 1 HCN Dusted all telescopes and UV cabinets

2:24 PM 1/9/2025 Montrose Remote KL BAD WEATHER

1:40 PM 1/20/2025 Montrose Remote KL BAD WEATHER 1/17-1/20 snow and cold

2:14 PM 1/21/2025 Montrose Onsite ML aligned TDLs

1:00 PM 1/29/25 MONTROSE ONSITE TC, CF ALIGNED PATH 1 HCN

12:15 PM 2/5/25 MONTROSE ONSITE CF ALIGNED PATH 1 HCN & H2S

10:55 AM 2/11/25 Montrose ONSITE EO
ALIGNED PATH 2 HCN, SNOW STARTED TO COME DOWN MORE & COULD NOT ALIGN PATH 1 HCN AND H2S.

11:50 AM 2/16/2025 Montrose Remote KL bad weather 2/14-2/16

2:12 PM 2/18/2025 Montrose Onsite ML Ozone filter change and realignment UV Path 1

3:21 PM 2/25/2025 Montrose Onsite TC Aligned H2S and HCN Path 1

5:10 PM 3/5/2025 Montrose Onsite JG, CN, ML Calibrated UVs and TDLs

12:48 PM 3/20/2025 Montrose onsite Aligned HCN path 1 Aligned UV Paths 1 & 2

11:45 AM 4/10/2025 TC Montrose onsite Aligned H2S ans HCN path 1

2:21 PM 4/15/2025 Montrose Onsite ML Cleaned TDL reflectors and aligned TDL paths 1 & 2 H2S and HCN

11:06 AM 5/14/2025 Montrose Onsite ML Checked alignment on H2S and HCN Path 1

12:54 PM 5/23/2025 Montrose Onsite ML

Aligned H2S and HCN path 1

11:23 AM 5/28/25 MONTROSE ONSITE CF ALIGNED HCN PATH 1 AFTER RETRO CLEANED

9:20 AM 1/8/2025 Montrose Remote KL Bad weather conditions 1/6-1/8

12:26 PM 1/9/2025 Montrose Remote KL BAD WEATHER

11:43 AM 1/20/2025 Montrose Remote KL BAD WEATHER: 1/17-1/20

12:10 PM 1/21/2025 Montrose Onsite ML aligned TDLs

10:16 AM 1/25/2025 Montrose Remote KL bad weather conditions 1/25-1/26

9:47 AM 2/16/2025 Montrose Remote kL bad weather 2/14-2/16

11:48 AM 2/18/2025 Montrose onsite ML Ozone filter change UV path 4 and realignment Retro filter change

12:35 PM 2/18/2025 Montrose Onsite ML Retro filter and UV path 3 alignment 2:08 PM 2/25/2025 Montrose Onsite TC Aligned H2S and HCN Path 4 Aligned UV Path 4 Started data back up at UV PATH 3

3:10 PM 3/5/2025 Montrose Onsite JG, CN, ML calibrated UVs and TDLs

11:41 PM 3/20/2025 Montrose Onsite ML Aligned UV path 4 Aligned TDL H2S & HCN Path 4

12:32 PM 4/15/2025 Montrose Onsite ML Clenaed TDL reflectos and checked alignment on TDL H2S and HCN for paths 3 & 4

2:00 PM 4/30/25 MONTROSE ONSITE CF ALIGNED UV PATH 4

11:19 AM 1/8/2025 Montrose Remote KL

Bad weather conditions 1/6-1/8

12:35 PM 1/8/25 Montrose onsite CF Aligned path 5 & 6 UVs, path 5 HCN & H@S, and path 6 H2S. Dusted all telescopes and UV cabinets

2:26 PM 1/9/2025 Montrose Remote KL bad weather 1:45 PM 1/20/2025 Montrose Remote KL BAD WEATHER: 1/17-1/20

2:05 PM 1/21/2025 Montrose Onsite ML aligned TDLs aligned UV

12:18 PM 1/25/2025 Montrose Remote KL bad weather 1/25-1/26

1:26 PM 1/29/2025 Montrose Onsite TC aligned path 5 TDLs/UV. aligned path 6 UV.

4:39 PM 2/1/2025 Montrose Onsite EO aligned tdls paths 5 and 6 for both H2S and HCN aligned uv path 6

11:25AM 2/5/25 MONTROSE ONSITE CF ALIGNED PATH 5 HCN ALIGNED PATH 6 UV

12:00 PM 2/11/25 MONTROSE ONSITE EO ALIGNED PATH 5 HCN ALIGNED PATH 6 HCN AND H2S

11:51 AM 2/16/2025 Montrose Remote KL Bad weather 2/14-2/16 10:40 AM 2/18/2025 Montrose Onsite ML UV Bulb change shelter 1 UV path 6

11:40 AM 2/18/2025 Montrose Onsite ML Ozone filter change UV path 5 and realignment

12:02 AM 2/18/2025 Montrose Onsite ML Ozone filter change UV pah 6 and relaignment

3:21 PM 2/25/2025 Montrose Onsite TC Aligned H2S and HCN Path 6

5:12 PM 3/5/2025 Montrose Onsite JG, CN, ML calibrated UVs and TDLs

11:55 AM 3/11/2025 Montrose Onsite EO DATData Backup for UV path 5 AliAligned H2S & HCN Paths 6 D. Appendix D: Non-Conformance/Corrective Action Data Sheets

Implementation Date: February 07, 2024 Form Owner (Department): MAQS Form Approval: AHeitmann

#### **Non-Conformance Report**

Project: PROJ-043815	Month: December 2024	
LOCATION/SITE: Goodrich Corporation in Pueblo	Parameter(s) Affected: Benzene Path 4	
Begin Date and Time (LST): 12/20/24 08:40AM       End Date and Time (LST): 12/20/24 1:35PM		
Equipment: UVDOAS Path 4	S/N#: N/A	
<b>Description of Malfunction or Problem:</b> Make specific reference to Assignable Cause(s). All tests results should be documented on appropriate form(s).		
UVDOAS computer went through updates and the so	tware could not restart.	
Investigative Actions: Describe Assignable Cause(s). M results. All tests results should be documented on approp	ake specific reference to all dates, times and performance test riate form(s).	
Updates on UVDOAS Path 4 computer. No benzene d	ata were collected during this time because the software	
was not on.		
Corrective Action Taken: Make specific reference to all	dates, times and performance test results.	
Remoted into the UVDOAS computer once the update	s were installed and manually restarted the software.	
Is Problem Fully Resolved? <b>Yes <u>x</u> No <u>If "NO</u> Report when problem is fully resolved)</b>	', Describe Further Action Required: (File updated NC/CA	
Additional Attachments or Information? Yes No	c Client Notified? Yes xIf so, date	
Field Operator's Assessment of Data Status: (Check One	) 🗆 Valid 🛛 Suspect 🛛 🖾 Invalid	
Additional notes on Data Validity Status: Benzene Path 4	data were not collected during this time.	
Originator's Signature:	Katia Liangou	

QA Review: <u>Aricia Boyd</u>

Implementation Date: February 07, 2024 Form Owner (Department): MAQS Form Approval: AHeitmann

#### **Non-Conformance Report**

ed: Benzene Path 4 Time (LST): 12/20/24 1:35Pl e Cause(s). All tests results sh art.	<b>M</b> hould be		
ed: Benzene Path 4 Time (LST): 12/20/24 1:35PI e Cause(s). All tests results sh art.	M hould be		
Time (LST): 12/20/24 1:35PI e Cause(s). All tests results sh art.	hould be		
e Cause(s). All tests results sh a <b>rt.</b>	hould be		
e Cause(s). All tests results sh a <b>rt.</b>	hould be		
art.			
to all dates, times and perform	mance test		
Updates on UVDOAS Path 4 computer. No benzene data were collected during this time because the software was not on.			
Corrective Action Taken: Make specific reference to all dates, times and performance test results. Remoted into the UVDOAS computer once the updates were installed and manually restarted the software.			
ion Required: (File updated N	IC/CA		
ion Required: (File updated N slf so, date	IC/CA		
ion Required: (File updated N sxlf so, date □ Suspect   ⊠	IC/CA  ⊴ Invalid		
t	tion Required: (File updated N ssxIf so, date		

Originator's Signature: Katia Liangou

QA Review: Aricia Boyd

1

Implementation Date: February 07, 2024 Form Owner (Department): MAQS Form Approval: AHeitmann

#### **Non-Conformance Report**

Project: PROJ-043815	Month: February 202	25	
	-		
LOCATION/SITE: Goodrich Corporation in Pueblo	Parameter(s) Affected: B	enzene Path 6	
Begin Date and Time (LST): 2/09/25 11:55PM	End Date and Time	e (LST): 2/10/25 7:4	40AM
Equipment: UVDOAS Path 6	S/N#: N/A		
<b>Description of Malfunction or Problem:</b> Make specific for documented on appropriate form(s).	eference to Assignable Cat	use(s). All tests rest	ults should be
UVDOAS computer went through updates and the so	tware could not restart.		
Investigative Actions: Describe Assignable Cause(s). M	ake specific reference to al	I dates, times and p	erformance test
results. All tests results should be documented on approp	riate form(s).	<i>,</i> , , , , , , , , , , , , , , , , , ,	
Updates on UVDOAS Path 6 computer. No benzene d	ata were collected during	this time because	the software
was not on.			
Corrective Action Taken: Make specific reference to all	dates, times and performan	nce test results.	
Remoted into the UVDOAS computer once the update	s were installed and man	ually restarted the	software.
Is Problem Fully Resolved? <b>Yes <u>x</u> No</b> If "NO", Describe Further Action Required: (File updated NC/CA Report when problem is fully resolved)			
Additional Attachments or Information? <b>Yes No</b>	Client Notified? Yes	<b>no<u>x</u></b> If so,	date
Field Operator's Assessment of Data Status: (Check One	) 🛛 🖓 Valid	□ Suspect	🛛 Invalid
Additional notes on Data Validity Status: Benzene Path 6 data were not collected during this time.			
Originator's Signature: _	Katia Liango	<u>ou</u>	
QA Revie	w: <u>Aricia Boy</u> g	d	

0

E. Appendix E: Calibration Verification Forms



Form Title: UVDOAS Calibration Form	Implementation Date: July 10, 2024
Document Number: 331AA-OPS-FM-13	Form Owner (Department): MAQS
Revision Number: Rev. 0	Form Approval: Katia Liangou
Operator Name(s): Katia Liangou	_ Test Date (YYYY/MM/DD): <u>3/5/25</u>
Instrument Model: UV Bi Path 1	_ Instrument Serial Number:
Instrument Parameters	

Optical Path Length (meters)	613 m/ 0.00235m
Maximum Intensity (%)	97
Integration Time (ms)	96

Standard I	nformation
Benzene Standard Concentration (PPM)	100

File #	Benzene Concentration (PPM)	Measured Concentration (PPM)	Error (%)
1	100	116	16
2	100	118	18
3	100	114	14
4	100	123	23
5	100	105	5
Averages	100	115	15.2



Form Title: UVDOAS Calibration Form	Implementation Date: July 10, 2024
Document Number: 331AA-OPS-FM-13	Form Owner (Department): MAQS
Revision Number: Rev. 0	Form Approval: Katia Liangou

	Calculated Values	Expected Values
Overall Percent Precision	93.4	≥ 75%
Overall Percent Error	15.2	≤ 30%

Notes:

Calibration verification passed.

Operator's Signature Katia Liangou Witness's Signature James Garrett



Form Title: UVDOAS Calibration Form	Implementation Date: July 10, 2024
Document Number: 331AA-OPS-FM-13	Form Owner (Department): MAQS
Revision Number: Rev. 0	Form Approval: Katia Liangou
Operator Name(s): Katia Liangou	_Test Date (YYYY/MM/DD): <u>3/5/25</u>
Instrument Model: UV Bi Path 1	_Instrument Serial Number:
Instrument Parameters	

Instrument Parameters	
Optical Path Length (meters)	613 m/ 0.0235m
Maximum Intensity (%)	97
Integration Time (ms)	95

Standard I	nformation
Benzene Standard Concentration (PPM)	200

File #	Benzene Concentration (PPM)	Measured Concentration (PPM)	Error (%)
1	200	183	8.5
2	200	217	8.5
3	200	191	4.5
4	200	193	3.5
5	200	193	3.5
Averages	200	195	5.7



Form Title: UVDOAS Calibration Form	Implementation Dat
Document Number: 331AA-OPS-FM-13	Form Owner (Depar
Revision Number: Rev. 0	Form Approval: Ka

te: July 10, 2024 rtment): MAQS itia Liangou

	Calculated Values	Expected Values
Overall Percent Precision	93.6	≥ 75%
Overall Percent Error	5.7	≤ 30%

Notes: Calibration verification passed.

Operator's Signature Katia Liangou Witness's Signature



Form Owner (Department): MAQS Form Approval: Katia Liangou
t Date (YYYY/MM/DD): <u>3/5/25</u> strument Serial Number:

Instrument Parameters		
Optical Path Length (meters)	566 m/ 0.047m	
Maximum Intensity (%)	83	
Integration Time (ms)	63	

Standard I	nformation
Benzene Standard Concentration (PPM)	100

File #	Benzene Concentration (PPM)	Measured Concentration (PPM)	Error (%)
1	100	95	5
2	100	97	3
3	100	97	3
4	100	91	9
5	100	110	10
Averages	100	98	6



Form Title: UVDOAS Calibration Form	Impleme
Document Number: 331AA-OPS-FM-13	Form Ow
Revision Number: Rev. 0	Form Ap

ntation Date: July 10, 2024 vner (Department): MAQS proval: Katia Liangou

	Calculated Values	Expected Values
Overall Percent Precision	92.9	≥ 75%
Overall Percent Error	6	≤ 30%

Notes:

Calibration verification passed.

Operator's Signature Katia Liangou Witness's Signature James Garrett

**Montrose Air Quality Services** 5120 Northshore Dr. North Little Rock, AR 72218 T: (501-)900-6400 www.montrose-env.com



Form Title: UVDOAS Calibration Form Document Number: 331AA-OPS-FM-13 Revision Number: Rev. 0	Implementation Date: July 10, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou
Operator Name(s): Katia Liangou	Test Date (YYYY/MM/DD): 3/5/25
Instrument Model: UV Mono Path 2	Instrument Serial Number:

Instrument Parameters	
Optical Path Length (meters)	566 m/ 0.047m
Maximum Intensity (%)	76
Integration Time (ms)	56

Standard I	nformation
Benzene Standard Concentration (PPM)	200

File #	Benzene Concentration (PPM)	Measured Concentration (PPM)	Error (%)
1	200	234	17
2	200	227	13.5
3	200	229	14.5
4	200	236	18
5	200	226	13
Averages	200	230	15.2



Form Title: UVDOAS Calibration Form	Implementation Date: July 10, 2024
Document Number: 331AA-OPS-FM-13	Form Owner (Department): MAQS
Revision Number: Rev. 0	Form Approval: Katia Liangou

	Calculated Values	Expected Values
Overall Percent Precision	97.8	≥ 75%
Overall Percent Error	15.2	≤ 30%

Notes:

Calibration verification passed.

Operator's Signature Katia Liangou Witness's Signature James Garrett

**Montrose Air Quality Services** 5120 Northshore Dr. North Little Rock, AR 72218 T: (501-)900-6400 www.montrose-env.com



Form Title: UVDOAS Calibration Form	n Implementation Date: July 10, 2024	
Document Number: 331AA-OPS-FM-	-13 Form Owner (Department): MAQS	
Revision Number: Rev. 0	Form Approval: Katia Liangou	
Operator Name(s): Katia Liangou	Test Date (YYYY/MM/DD): <u>3/5/25</u>	
Instrument Model: UV Bi Path 3	Instrument Serial Number:	
Instrument Parameters		

Instrument Parameters	
Optical Path Length (meters)	613 m/ 0.0235m
Maximum Intensity (%)	95.2
Integration Time (ms)	38

Standard I	nformation
Benzene Standard Concentration (PPM)	100

File #	Benzene Concentration (PPM)	Measured Concentration (PPM)	Error (%)
1	100	95	5
2	100	92	8
3	100	83	17
4	100	82	18
5	100	90	10
Averages	100	88.4	11.6



Form Title: UVDOAS Calibration Form	Implementation Date: July 10, 2024
Document Number: 331AA-OPS-FM-13	Form Owner (Department): MAQS
Revision Number: Rev. 0	Form Approval: Katia Liangou

	Calculated Values	Expected Values
Overall Percent Precision	94.3	≥ 75%
Overall Percent Error	11.6	≤ 30%

Notes:

Calibration verification passed.

Operator's Signature Katia Liangou Witness's Signature James Garrett

**Montrose Air Quality Services** 5120 Northshore Dr. North Little Rock, AR 72218 T: (501-)900-6400 www.montrose-env.com



Form Title: UVDOAS Calibration Form Document Number: 331AA-OPS-FM-13 Revision Number: Rev. 0	Implementation Date: July 10, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou
Operator Name(s): Katia Liangou	_Test Date (YYYY/MM/DD): <u>3/5/25</u>
Instrument Model: UV Bi Path 3	_ Instrument Serial Number:

Instrument Parameters		
Optical Path Length (meters)	613 m/ 0.0235m	
Maximum Intensity (%)	93	
Integration Time (ms) 38		

Standard I	nformation
Benzene Standard Concentration (PPM)	200

File #	Benzene Concentration (PPM)	Measured Concentration (PPM)	Error (%)
1	200	207	3.5
2	200	190	5
3	200	212	6
4	200	219	9.5
5	200	209	4.5
Averages	200	207	5.7



Form Title: UVDOAS Calibration Form	Implementation D
Document Number: 331AA-OPS-FM-13	Form Owner (Dep
Revision Number: Rev. 0	Form Approval:

ate: July 10, 2024 artment): MAQS Katia Liangou

	Calculated Values	Expected Values
Overall Percent Precision	94.6	≥ 75%
Overall Percent Error	5.7	≤ 30%

Notes:

Calibration verification passed.



Form Title: UVDOAS Calibration Form Document Number: 331AA-OPS-FM-13 Revision Number: Rev. 0	Implementation Date: July 10, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou
Operator Name(s): Katia Liangou Instrument Model: UV Bi Path 4	_ Test Date (YYYY/MM/DD): <u>12/30/2024</u> _ Instrument Serial Number:
Ins	strument Parameters

Instrument Parameters		
Optical Path Length (meters)	566 m/ 0.0235m	
Maximum Intensity (%)	96	
Integration Time (ms) 95		

Standard I	nformation
Benzene Standard Concentration (PPM)	100

File #	Benzene Concentration (PPM)	Measured Concentration (PPM)	Error (%)
1	100	124	24
2	100	113	13
3	100	129	29
4	100	127	27
5	100	120	20
Averages	100	123	22.6



Form Title: UVDOAS Calibration Form	Implementation Date: July 10, 2024
Document Number: 331AA-OPS-FM-13	Form Owner (Department): MAQS
Revision Number: Rev. 0	Form Approval: Katia Liangou

	Calculated Values	Expected Values
Overall Percent Precision	93.7	≥ 75%
Overall Percent Error	22.6	≤ 30%

Notes:

Calibration verification passed.



Form Title: UVDOAS Calibration Form Document Number: 331AA-OPS-FM-13 Revision Number: Rev. 0	Implementation Date: July 10, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou
Operator Name(s): Katia Liangou	_Test Date (YYYY/MM/DD): <u>3/5/2025</u>
Instrument Model: UV BI Path 4	_ Instrument Serial Number:

Instrument Parameters	
Optical Path Length (meters)	566 m/ 0.0235m
Maximum Intensity (%)	96
Integration Time (ms) 95	

Standard I	nformation
Benzene Standard Concentration (PPM)	200

File #	Benzene Concentration (PPM)	Measured Concentration (PPM)	Error (%)
1	200	197	1.5
2	200	207	3.5
3	200	167	16.5
4	200	204	2
5	200	219	9.5
Averages	200	199	6.6



Form Title: UVDOAS Calibration Form	Implementation Dat
Document Number: 331AA-OPS-FM-13	Form Owner (Depa
Revision Number: Rev. 0	Form Approval: Ka

te: July 10, 2024 rtment): MAQS atia Liangou

	Calculated Values	Expected Values
Overall Percent Precision	90.3	≥ 75%
Overall Percent Error	6.6	≤ 30%

Notes:

Calibration verification passed.

**Montrose Air Quality Services** 5120 Northshore Dr. North Little Rock, AR 72218 T: (501-)900-6400 www.montrose-env.com



Form Title: UVDOAS Calibration Form Document Number: 331AA-OPS-FM-13 Revision Number: Rev. 0	Implementation Date: July 10, 2024           Form Owner (Department): MAQS           Form Approval: Katia Liangou
Operator Name(s): Katia Liangou	Test Date (YYYY/MM/DD): <u>3/5/25</u>
Instrument Model: UV Mono Path 5	Instrument Serial Number:

Instrument Parameters	
Optical Path Length (meters) 594 m/ 0.047m	
Maximum Intensity (%)	97
Integration Time (ms) 91	

Standard I	nformation
Benzene Standard Concentration (PPM)	100

File #	Benzene Concentration (PPM)	Measured Concentration (PPM)	Error (%)
1	100	112	12
2	100	135	35
3	100	100	0
4	100	132	32
5	100	112	12
Averages	100	118	18.2



Form Title: UVDOAS Calibration Form	Implementation Date: July 10, 2024
Document Number: 331AA-OPS-FM-13	Form Owner (Department): MAQS
Revision Number: Rev. 0	Form Approval: Katia Liangou

	Calculated Values	Expected Values
Overall Percent Precision	85.2	≥ 75%
Overall Percent Error	18.2	≤ 30%

Notes:

Calibration verification passed.

**Montrose Air Quality Services** 5120 Northshore Dr. North Little Rock, AR 72218 T: (501-)900-6400 www.montrose-env.com



٦

Form Title: UVDOAS Calibration Form Document Number: 331AA-OPS-FM-13 Revision Number: Rev. 0	Implementation Date: July 10, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou
Operator Name(s): Katia Liangou	_Test Date (YYYY/MM/DD): <u>3/5/25</u>
Instrument Model: UV Mono Path 5	_ Instrument Serial Number:

Instrument Parameters	
Optical Path Length (meters) 594 m/ 0.047m	
Maximum Intensity (%)	96
Integration Time (ms) 90	

Standard I	nformation
Benzene Standard Concentration (PPM)	200

File #	Benzene Concentration (PPM)	Measured Concentration (PPM)	Error (%)
1	200	206	3
2	200	217	8.5
3	200	200	0
4	200	214	7
5	200	210	5
Averages	200	209	4.7



Form Title: UVDOAS Calibration Form	Implementation Date: July 10, 2024
Document Number: 331AA-OPS-FM-13	Form Owner (Department): MAQS
Revision Number: Rev. 0	Form Approval: Katia Liangou

	Calculated Values	Expected Values
Overall Percent Precision	96.7	≥ 75%
Overall Percent Error	4.7	≤ 30%

Notes:

Calibration verification passed.

Operator's Signature Katia Liangou Witness's Signature



Form Title: UVDOAS Calibration Form	Implementation Date: July 10, 2024
Document Number: 331AA-OPS-FM-13	Form Owner (Department): MAQS
Revision Number: Rev. 0	Form Approval: Katia Liangou
Operator Name(s): Katia Liangou	_Test Date (YYYY/MM/DD): <u>3/5/25</u>
Instrument Model: UV Bi Path 6	_Instrument Serial Number:

Instrument Parameters	
Optical Path Length (meters)	569 m/ 0.0235m
Maximum Intensity (%)	88
Integration Time (ms) 60	

Standard I	nformation
Benzene Standard Concentration (PPM)	100

File #	Benzene Concentration (PPM)	Measured Concentration (PPM)	Error (%)
1	100	113	13
2	100	111	11
3	100	99	1
4	100	133	33
5	100	117	17
Averages	100	115	15



Form Title: UVDOAS Calibration Form	Implementation Date: July 10, 2024
Document Number: 331AA-OPS-FM-13	Form Owner (Department): MAQS
Revision Number: Rev. 0	Form Approval: Katia Liangou

	Calculated Values	Expected Values
Overall Percent Precision	87.7	≥ 75%
Overall Percent Error	15	≤ 30%

Notes:

Calibration verification passed.

**Montrose Air Quality Services** 5120 Northshore Dr. North Little Rock, AR 72218 T: (501-)900-6400 www.montrose-env.com



Form Title: UVDOAS Calibration Form Document Number: 331AA-OPS-FM-13 Revision Number: Rev. 0	Implementation Date: July 10, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou
Operator Name(s): Katia Liangou	_Test Date (YYYY/MM/DD): <u>3/5/25</u>
Instrument Model: UV BI Path 6	_ Instrument Serial Number:

Instrument Parameters		
Optical Path Length (meters) 569 m/ 0.0235m		
Maximum Intensity (%)	81	
Integration Time (ms) 36		

Standard I	nformation
Benzene Standard Concentration (PPM)	200

File #	Benzene Concentration (PPM)	Measured Concentration (PPM)	Error (%)
1	200	214	7
2	200	194	3
3	200	248	24
4	200	237	18.5
5	200	183	8.5
Averages	200	215	12.2



Form Title: UVDOAS Calibration Form	Implementation Date: July 10, 2024	
Document Number: 331AA-OPS-FM-13	Form Owner (Department): MAQS	
Revision Number: Rev. 0	Form Approval: Katia Liangou	

	Calculated Values	Expected Values
Overall Percent Precision	86.2	≥ 75%
Overall Percent Error	12.2	≤ 30%

Notes:

Calibration verification passed.

Operator's Signature Katia Liangou Witness's Signature



Г

# Page 1 of 2 TDL Calibration Form

Form Title: TDL Calibration Form	Implementation Date: August 8, 2024	
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS	
Revision Number: Rev. 1	Form Approval: Katia Liangou	

\_\_\_\_\_

Operator Name(s): Katia Liangou	Test Date (YYYY/MM/DD): 3/5/25
---------------------------------	--------------------------------

Instrument Model: H2S Path 1 Instrument Serial Number: \_\_\_\_\_

Instrument Parameters		
Optical Path separation(meters-one-way)	558 m	
Compound (H2S/HCN)	H2S	

Standard Information		
Compound External Audit Cell Concentration (PPMM)	500 PPMM	

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	500	526	5.2
2	500	560	12
3	500	536	7.2
4	500	534	6.8
5	500	558	11.6
Averages	500	543	8.6

	Calculated Values	Expected Values
Overall Percent Precision	96.9%	≥ 80%
Overall Percent Error	8.6%	≤ 30%



#### Page 2 of 2 **TDL Calibration Form**

Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1

Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

\_Witness Signature(s): <u>James Garrett</u>

**Montrose Air Quality Services** 5120 Northshore Dr. North Little Rock, AR 72218 T: (501-)900-6400 www.montrose-env.com


Г

## Page 1 of 2 TDL Calibration Form

Form Title: TDL Calibration Form	Implementation Date: August 8, 2024
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS
Revision Number: Rev. 1	Form Approval: Katia Liangou

\_\_\_\_\_

\_\_\_\_\_

Operator Name(s): Katia Liangou	Test Date (YYYY/MM/DD): 3/5/2025
---------------------------------	----------------------------------

Instrument Model: H2S Path 1 Instrument Serial Number:

Instrument Parameters		
Optical Path separation(meters-one-way)	558 m	
Compound (H2S/HCN)	H2S	

Standard I	nformation
Compound External Audit Cell Concentration (PPMM)	625 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	625	642	2.7
2	625	680	8.8
3	625	610	2.4
4	625	624	0.2
5	625	622	0.5
Averages	625	636	2.9

	Calculated Values	Expected Values
Overall Percent Precision	95.6%	≥ 80%
Overall Percent Error	2.9 %	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

\_Witness Signature(s):

James Garrett



Form Title: TDL Calibration Form	Implementation Date: August 8, 2024
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS
Revision Number: Rev. 1	Form Approval: Katia Liangou

\_\_\_\_

	Operator Name(s): Katia Liangou	Test Date (YYYY/MM/DD): 3/5/2025
--	---------------------------------	----------------------------------

Instrument Model: H2S Path 2 Instrument Serial Number:

Instrument Parameters		
Optical Path separation(meters-one-way)	283 m	
Compound (H2S/HCN)	H2S	

Standard I	nformation
Compound External Audit Cell Concentration (PPMM)	500 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	500	492	1.6
2	500	496	0.8
3	500	490	2
4	500	484	3.2
5	500	472	5.6
Averages	500	487	2.6

	Calculated Values	Expected Values
Overall Percent Precision	98.1%	≥ 80%
Overall Percent Error	2.6%	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

\_Witness Signature(s):

James Garrett



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1	Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou	
Operator Name(s): Katia Liangou Te Instrument Model: H2S Path 2 Inst	est Date (YYYY/MM/DD): <u>3/5/2025</u> strument Serial Number:	
Instrument Parameters		
Optical Path separation(meters-one-way)	283 m	
Compound (H2S/HCN)	H2S	

Standard I	nformation
Compound External Audit Cell Concentration (PPMM)	625 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	625	704	12.6
2	625	704	12.6
3	625	702	12.3
4	625	718	14.9
5	625	614	1.8
Averages	625	688	10.8

	Calculated Values	Expected Values
Overall Percent Precision	93.2	≥ 80%
Overall Percent Error	10.8	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

James Garrett \_Witness Signature(s):



Form Title: TDL Calibration Form	Implementation Date: August 8, 2024
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS
Revision Number: Rev. 1	Form Approval: Katia Liangou
Operator Name(s): Katia Liangou	_Test Date (YYYY/MM/DD): <u>3/5/25</u>

Instrument Model: H2S Path 3 Instrument Serial Number: \_\_\_\_\_

Instrument Parameters	
Optical Path separation(meters-one-way) 613 m	
Compound (H2S/HCN)	H2S

Standard I	nformation
Compound External Audit Cell Concentration (PPMM)	500 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	500	484	3.2
2	500	478	4.4
3	500	466	6.8
4	500	468	6.4
5	500	478	4.4
Averages	500	475	5

	Calculated Values	Expected Values
Overall Percent Precision	98.5%	≥ 80%
Overall Percent Error	5%	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

\_Witness Signature(s): <u>James</u> Jarrett



Form Title: TDL Calibration Form	Implementation Date: August 8, 2024
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS
Revision Number: Rev. 1	Form Approval: Katia Liangou
Operator Name(s): Katia Liangou	_Test Date (YYYY/MM/DD): <u>3/5/2025</u>

Instrument Model: H2S Path 3 \_\_\_\_\_ Instrument Serial Number: \_\_\_\_\_\_

Instrument Parameters	
Optical Path separation(meters-one-way) 613 m	
Compound (H2S/HCN)	H2S

Standard I	nformation
Compound External Audit Cell Concentration (PPMM)	625 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	625	586	6.2
2	625	570	8.8
3	625	584	6.6
4	625	584	6.6
5	625	584	6.6
Averages	625	582	6.9

	Calculated Values	Expected Values
Overall Percent Precision	99%	≥ 80%
Overall Percent Error	6.9 %	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):



Form Title: TDL Calibration Form	Implementation Date: August 8, 2024
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS
Revision Number: Rev. 1	Form Approval: Katia Liangou

\_\_\_\_\_

Operator Name(s):	Katia Liangou	Test Date (	(YYYY/MM/DD): 3/5/25
	<u>~</u>		

Instrument Model: H2S Path 4 Instrument Serial Number: \_\_\_\_\_

Instrument Parameters	
Optical Path separation(meters-one-way) 566 m	
Compound (H2S/HCN)	H2S

Standard I	nformation
Compound External Audit Cell Concentration (PPMM)	500 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	500	418	16.4
2	500	398	20.4
3	500	402	19.6
4	500	390	22
5	500	396	20.8
Averages	500	401	19.8

	Calculated Values	Expected Values
Overall Percent Precision	97.9%	≥ 80%
Overall Percent Error	19.8%	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

\_Witness Signature(s):

James Garrett



Form Title: TDL Calibration Form	Implementation Date: August 8, 2024
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS
Revision Number: Rev. 1	Form Approval: Katia Liangou

-----

Operator Name(s):	Katia Liangou	Test Date (	(YYYY/MM/DD): 3/5/25
	<u>~</u>		

Instrument Model: H2S Path 4 Instrument Serial Number: \_\_\_\_\_

Instrument Parameters		
Optical Path separation(meters-one-way) 566 m		
Compound (H2S/HCN)	H2S	

Standard Information	
Compound External Audit Cell Concentration (PPMM)	625 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	625	552	11.7
2	625	556	11
3	625	560	10.4
4	625	568	9.1
5	625	546	12.6
Averages	625	556	11

	Calculated Values	Expected Values
Overall Percent Precision	98.7%	≥ 80%
Overall Percent Error	11%	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

\_Witness Signature(s):

James Garrett



Г

# Page 1 of 2 TDL Calibration Form

Form Title: TDL Calibration Form	Implementation Date: August 8, 2024
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAOS
Revision Number: Rev. 1	Form Approval: Katia Liangou
Operator Name (s), Katia Liangou	Test Date ()////////////////////////////////////

\_\_\_\_\_

Operator Name(s):	Katia Liangou	Test Date (YYYY/MM/DD): 3/5/	-
-------------------	---------------	------------------------------	---

Instrument Model: H2S Path 5 \_\_\_\_\_ Instrument Serial Number: \_\_\_\_\_\_

Instrument Parameters	
Optical Path separation(meters-one-way)	297 m
Compound (H2S/HCN)	H2S

Standard I	nformation
Compound External Audit Cell Concentration (PPMM)	500 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	500	506	12
2	500	526	5.2
3	500	504	0.8
4	500	528	5.6
5	500	518	3.6
Averages	500	516	3.3

	Calculated Values	Expected Values
Overall Percent Precision	97.8%	≥ 80%
Overall Percent Error	3.3%	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):



Form Title: TDL Calibration Form	Implementation Date: August 8, 2024
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS
Revision Number: Rev. 1	Form Approval: Katia Liangou
Operator Name(s): Katia Liangou	_Test Date (YYYY/MM/DD): <u>3/5/25</u>

Operator Name(s): Katia Liangou Test	Date (YYYY/MM/DD): 3/5/25
--------------------------------------	---------------------------

Instrument Model: H2S Path 5 \_\_\_\_\_ Instrument Serial Number: \_\_\_\_\_\_

Instrument Parameters	
Optical Path separation(meters-one-way) 297 m	
Compound (H2S/HCN)	H2S

Standard I	nformation
Compound External Audit Cell Concentration (PPMM)	625 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	625	708	13.3
2	625	722	15.5
3	625	716	14.6
4	625	702	12.3
5	625	706	13
Averages	625	711	13.7

	Calculated Values	Expected Values
Overall Percent Precision	98.7%	≥ 80%
Overall Percent Error	13.7%	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

\_Witness Signature(s):

James Garrett



Г

## Page 1 of 2 TDL Calibration Form

Form Title: TDL Calibration Form	Implementation Date: August 8, 2024
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS
Revision Number: Rev. 1	Form Approval: Katia Liangou

\_\_\_\_\_

Operator Name(s):	Katia Liangou	Test Date (	(YYYY/MM/DD): 3/5/25
	<u>~</u>		

Instrument Model: H2S Path 6 Instrument Serial Number: \_\_\_\_\_

Instrument Parameters		
Optical Path separation(meters-one-way) 569 m		
Compound (H2S/HCN)	H2S	

Standard I	nformation
Compound External Audit Cell Concentration (PPMM)	500 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	500	528	5.6
2	500	514	2.8
3	500	522	4.4
4	500	524	4.8
5	500	518	3.6
Averages	500	521	4.2

	Calculated Values	Expected Values
Overall Percent Precision	98.9%	≥ 80%
Overall Percent Error	4.2%	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

\_Witness Signature(s): <u>James Garrett</u>



Г

## Page 1 of 2 TDL Calibration Form

Form Title: TDL Calibration Form	Implementation Date: August 8, 2024
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS
Revision Number: Rev. 1	Form Approval: Katia Liangou

\_\_\_\_

Operator Name(s):	Katia Liangou	Test Date (	(YYYY/MM/DD): 3/5/25
	<u>~</u>		

Instrument Model: H2S Path 6 Instrument Serial Number:

Instrument Parameters	
Optical Path separation(meters-one-way)	569 m
Compound (H2S/HCN)	H2S

Standard I	nformation
Compound External Audit Cell Concentration (PPMM)	625 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	625	716	14.6
2	625	682	9.1
3	625	682	9.1
4	625	682	9.1
5	625	682	9.1
Averages	625	689	10.2

	Calculated Values	Expected Values
Overall Percent Precision	97.6%	≥ 80%
Overall Percent Error	10.2%	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

\_Witness Signature(s): <u>James</u> Garrett



Form Title: TDL Calibration Form	Implementation Date: August 8, 2024	
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS	
Revision Number: Rev. 1	Form Approval: Katia Liangou	
Operator Name(s): Katia Liangou	_Test Date (YYYY/MM/DD): <u>3/5/25</u>	

Operator Name(s): Ka	tia Liangou	Test Date (YYYY/MM/DD):	3/5/	2
----------------------	-------------	-------------------------	------	---

Instrument Model: HCN Path 1 Instrument Serial Number: \_\_\_\_\_

Instrument Parameters	
Optical Path separation(meters-one-way)	558 m
Compound (H2S/HCN)	HCN

Standard I	nformation
Compound External Audit Cell Concentration (PPMM)	420 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	420	496	18.1
2	420	498	18.6
3	420	494	17.6
4	420	492	17.1
5	420	492	17.1
Averages	420	494	17.7

	Calculated Values	Expected Values
Overall Percent Precision	99.4%	≥ 80%
Overall Percent Error	17.7 %	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

\_Witness Signature(s):

James Garrett



Г

# Page 1 of 2 TDL Calibration Form

Form Title: TDL Calibration Form	Implementation Date: August 8, 2024
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS
Revision Number: Rev. 1	Form Approval: Katia Liangou
Operator Name(s): Katia Liangou	_Test Date (YYYY/MM/DD): <u>3/5/25</u>

Operator Name(s):	Katia Liangou	_Test Date (YYYY/MM/DD): <u>3/5/2</u>	2
-------------------	---------------	---------------------------------------	---

Instrument Model: HCN Path 1 Instrument Serial Number: \_\_\_\_\_

Instrument Parameters	
Optical Path separation(meters-one-way) 558 m	
Compound (H2S/HCN)	HCN

Standard I	nformation
Compound External Audit Cell Concentration (PPMM)	1010 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	1010	1052	4.2
2	1010	1056	4.6
3	1010	1052	4.2
4	1010	1052	4.2
5	1010	1056	4.6
Averages	1010	1054	4.3

	Calculated Values	Expected Values
Overall Percent Precision	99.8%	≥ 80%
Overall Percent Error	4.3 %	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

\_Witness Signature(s):

James Garrett



Г

# Page 1 of 2 TDL Calibration Form

Form Title: TDL Calibration Form	Implementation Date: August 8, 2024
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS
Revision Number: Rev. 1	Form Approval: Katia Liangou
Operator Name(s). Katia Liangou	Test Date (YYYY/MM/DD): 3/5/25

\_\_\_\_\_

Operator Name(s):	Katia Liangou	Test Date (YYYY/MM/DD): 3/5/2
		- · · · · · · · · · · · · · · · · · · ·

Instrument Model: HCN Path 2 Instrument Serial Number: \_\_\_\_\_

Instrument Parameters	
Optical Path separation(meters-one-way) 283 m	
Compound (H2S/HCN)	HCN

Standard I	nformation
Compound External Audit Cell Concentration (PPMM)	420 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	420	492	17.1
2	420	490	16.7
3	420	490	16.7
4	420	492	17.1
5	420	492	17.1
Averages	420	491	17

	Calculated Values	Expected Values
Overall Percent Precision	99.7%	≥ 80%
Overall Percent Error	17%	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

\_Witness Signature(s):

James Garrett



Form Title: TDL Calibration Form	Implementation Date: August 8, 2024
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS
Revision Number: Rev. 1	Form Approval: Katia Liangou

\_\_\_\_\_

	3/5/25
--	--------

Instrument Model: HCN Path 2 Instrument Serial Number: \_\_\_\_\_

Instrument Parameters	
Optical Path separation(meters-one-way) 283 m	
Compound (H2S/HCN)	HCN

Standard I	nformation
Compound External Audit Cell Concentration (PPMM)	1010 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	1010	1054	4.4
2	1010	1054	4.4
3	1010	1054	4.4
4	1010	1054	4.4
5	1010	1054	4.4
Averages	1010	1054	4.4

	Calculated Values	Expected Values
Overall Percent Precision	100%	≥ 80%
Overall Percent Error	4.4%	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

\_Witness Signature(s):

James Garrett



Form Title: TDL Calibration Form	Implementation Date: August 8, 2024
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS
Revision Number: Rev. 1	Form Approval: Katia Liangou
Operator Name(s): Katia Liangou	_Test Date (YYYY/MM/DD): <u>3/5/25</u>

Operator Name(s): <u>Natia Liangou</u> Test Date	e (YYYY/MM/DD): 0/0/	~
--	----------------------	---

Instrument Model: HCN Path 3 Instrument Serial Number: \_\_\_\_

Instrument Parameters	
Optical Path separation(meters-one-way)	613 m
Compound (H2S/HCN)	HCN

Standard I	nformation
Compound External Audit Cell Concentration (PPMM)	420 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	420	484	15.2
2	420	482	14.8
3	420	482	14.8
4	420	480	14.3
5	420	480	14.3
Averages	420	482	14.7

	Calculated Values	Expected Values
Overall Percent Precision	99.6%	≥ 80%
Overall Percent Error	14.7%	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

\_Witness Signature(s): <u>James Jarrett</u>



Form Title: TDL Calibration Form	Implementation Date: August 8, 2024
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS
Revision Number: Rev. 1	Form Approval: Katia Liangou
Operator Name(s): Katia Liangou	_Test Date (YYYY/MM/DD): 3/5/25

Operator Name(s):	Katia Liangou	Test Date (YYYY/MM/DD):	3/5/2
		· · · · ·	

Instrument Model: HCN Path 3 Instrument Serial Number: \_\_\_\_\_

Instrument Parameters	
Optical Path separation(meters-one-way)	613 m
Compound (H2S/HCN)	HCN

Standard Information	
Compound External Audit Cell Concentration (PPMM)	1010 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	1010	1042	3.2
2	1010	1040	3
3	1010	1040	3
4	1010	1042	3.2
5	1010	1040	3
Averages	1010	1041	3

	Calculated Values	Expected Values
Overall Percent Precision	99.9%	≥ 80%
Overall Percent Error	3%	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

\_Witness Signature(s):

James Garrett



г

## Page 1 of 2 TDL Calibration Form

Form Title: TDL Calibration Form	Implementation Date: August 8, 2024	
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS	
Revision Number: Rev. 1	Form Approval: Katia Liangou	

\_\_\_\_\_

Operator Name(s): Katla Llangou Test Da	te (YYYY/MM/DD): 3/5/25
---	-------------------------

Instrument Model: HCN Path 4 Instrument Serial Number: \_\_\_\_\_

Instrument Parameters		
Optical Path separation(meters-one-way)	566 m	
Compound (H2S/HCN)	HCN	

Standard I	nformation
Compound External Audit Cell Concentration (PPMM)	420 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	420	418	0.4
2	420	398	5.2
3	420	402	4.2
4	420	390	7.1
5	420	396	5.7
Averages	420	401	4.6

	Calculated Values	Expected Values
Overall Percent Precision	97.5%	≥ 80%
Overall Percent Error	4.6%	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1

Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

\_Witness Signature(s): <u>James Garrett</u>


Form Title: TDL Calibration Form	Implementation Date: August 8, 2024
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS
Revision Number: Rev. 1	Form Approval: Katia Liangou

-----

Operator Name(s):	Katia Liangou	Test Date (	(YYYY/MM/DD): 3/5/25
	<u>~</u>		

Instrument Model: HCN Path 4 Instrument Serial Number: \_\_\_\_\_

Instrument Parameters		
Optical Path separation(meters-one-way) 566 m		
Compound (H2S/HCN)	HCN	

Standard I	nformation
Compound External Audit Cell Concentration (PPMM)	1010 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	1010	1036	2.6
2	1010	1038	2.8
3	1010	1040	3
4	1010	1034	2.4
5	1010	1038	2.8
Averages	1010	1037	2.7

	Calculated Values	Expected Values
Overall Percent Precision	99.8%	≥ 80%
Overall Percent Error	2.7%	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

\_Witness Signature(s): \_\_\_\_\_

James Garrett



Form Title: TDL Calibration Form	Implementation Date: August 8, 2024
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS
Revision Number: Rev. 1	Form Approval: Katia Liangou

\_\_\_\_\_

Operator Name(s): Katia Liangou	Test Date (YYYY/MM/DD): 3/5/25

Instrument Model: HCN Path 5 Instrument Serial Number:

Instrument Parameters		
Optical Path separation(meters-one-way) 297 m		
Compound (H2S/HCN)	HCN	

Standard I	nformation
Compound External Audit Cell Concentration (PPMM)	420 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	420	472	12.4
2	420	470	11.9
3	420	470	11.9
4	420	470	11.9
5	420	472	12.4
Averages	420	471	12.1

	Calculated Values	Expected Values
Overall Percent Precision	99.7%	≥ 80%
Overall Percent Error	12.1%	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

\_Witness Signature(s): <u>James Garrett</u>



Г

# Page 1 of 2 TDL Calibration Form

Form Title: TDL Calibration Form	Implementation Date: August 8, 2024
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS
Revision Number: Rev. 1	Form Approval: Katia Liangou

\_\_\_\_\_

Operator Name(s): Katia Liangou Test	Date (YYYY/MM/DD): 3/5/25
--------------------------------------	---------------------------

Instrument Model: HCN Path 5 Instrument Serial Number: \_\_\_\_\_

Instrument Parameters			
Optical Path separation(meters-one-way) 297 m			
Compound (H2S/HCN)	HCN		

Standard I	nformation
Compound External Audit Cell Concentration (PPMM)	1010 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	1010	1034	2.4
2	1010	1034	2.4
3	1010	1032	2.2
4	1010	1034	2.4
5	1010	1036	2.6
Averages	1010	1034	2.4

	Calculated Values	Expected Values
Overall Percent Precision	99.9%	≥ 80%
Overall Percent Error	2.4%	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

\_Witness Signature(s):

James Garrett



Г

# Page 1 of 2 TDL Calibration Form

Form Title: TDL Calibration Form	Implementation Date: August 8, 2024	
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS	
Revision Number: Rev. 1	Form Approval: Katia Liangou	

\_\_\_\_\_

Operator Name(s):	Katia Liangou	Test Date (	(YYYY/MM/DD): 3/5/25
	<u>~</u>		

Instrument Model: HCN Path 6 Instrument Serial Number: \_\_\_\_\_

Instrument Parameters			
Optical Path separation(meters-one-way) 569 m			
Compound (H2S/HCN)	HCN		

Standard I	nformation
Compound External Audit Cell Concentration (PPMM)	420 PPMM

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	420	484	15.2
2	420	484	15.2
3	420	482	14.8
4	420	482	14.8
5	420	482	14.8
Averages	420	483	15

	Calculated Values	Expected Values
Overall Percent Precision	99.7%	≥ 80%
Overall Percent Error	15%	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):

\_Witness Signature(s):

James Garrett



Г

# Page 1 of 2 TDL Calibration Form

Form Title: TDL Calibration Form	Implementation Date: August 8, 2024
Document Number: 331AA-OPS-FM-15	Form Owner (Department): MAQS
Revision Number: Rev. 1	Form Approval: Katia Liangou

\_\_\_\_\_

Operator Name(s):	Katia Liangou	Test Date (	(YYYY/MM/DD): 3/5/25
	<u>~</u>		

Instrument Model: HCN Path 6 Instrument Serial Number: \_\_\_\_\_

Instrument Parameters		
Optical Path separation(meters-one-way)	569 m	
Compound (H2S/HCN)	HCN	

Standard Information		
Compound External Audit Cell Concentration (PPMM)	1010 PPMM	

File #	Compound Concentration (PPMM)	Measured Concentration (PPMM)	Error (% Reading)
1	1010	1042	3.2
2	1010	1042	3.2
3	1010	1044	3.4
4	1010	1042	3.2
5	1010	1042	3.2
Averages	1010	1042	3.2

	Calculated Values	Expected Values
Overall Percent Precision	99.9%	≥ 80%
Overall Percent Error	3.2 %	≤ 30%



Form Title: TDL Calibration Form Document Number: 331AA-OPS-FM-15 Revision Number: Rev. 1 Implementation Date: August 8, 2024 Form Owner (Department): MAQS Form Approval: Katia Liangou

Notes:

Calibration verification passed.

Operator Signature(s):